

**Information Letter Number:** 242301**Date:** 03.06.2024**Title:** SFM3019 – datasheet modification**Product Identification:**

Order number	Material description
3.000.400	SFM3019

**Reason for Change:** Design Production Logistics Manufacturing Location Quality/Reliability Upgrade**Change Description:**

The mass flow meter SFM3019 datasheet has been corrected and updated. Relevant to the sensor's specifications are the following changes:

- Mention that the specifications are valid in dry air (footnote 1 page 2).
- Correction of the "max." noise specification (footnote 10 page 2).

Kindly find the details in the updated version of the datasheet released together with this infoletter.

**Identification Method to Distinguish Change:**

This letter informs about a correction of the datasheet but not a change of the sensor. Therefore, no distinction method is needed on the sensor.

**Samples:** available will be available in WW\_\_ not applicable**Quantifiable Impact on Quality & Reliability:**

No effect on form, fit or function.

**Estimated Implementation Date:** immediate**Sensirion Contact:** [PCN@sensirion.com](mailto:PCN@sensirion.com)

If you have questions with regard to this Information Letter, please send them to the Sensirion contact e-mail address listed above.